



This document certifies that the following product and specific batch has been manufactured to and complies with the company's standard Specification.


It further certifies that the batch has been tested by a diagnostic laboratory and complies with the minimum specification for this product.

PRODUCT CODE	RRSP78		
DESCRIPTION	Leishman Stain		
BATCH NO	Allocated at time of production		
Manufacture Date	3 Year Shelf Life	Expiry Date	3 Year Shelf Life

SPECIFICATION	A solution of Leishman stain (0.15% w/v) in Methyl alcohol (Methanol) for use in staining blood smears.		
 	<u>Contents:</u> Methanol: >99%		
<div style="border: 1px solid red; padding: 5px; font-size: small;"> This product has been registered as a medical devices with the "Medical and Healthcare Products Regulatory Agency", 11th Floor, Hannibal House, Elephant and Castle, London SE1 6TQ, UK under approval number IVD000444 which lists all EDMA Classifications applicable. </div>	Recommended dilution 1:1 or 1:2 in Sorensen phosphate buffer pH 6.5-6.8		

TESTING PROTOCOL <small>(WHERE APPLICABLE)</small>	This product (diluted 1:1) is tested in the staining of peripheral blood smears using our standard protocol.		
--	--	--	--

We confirm that this batch has been tested in our Laboratory and has been found to stain to our minimum standards and that sample will be retained for the life of the product



Peter Keenan
Commercial Director

Disclaimer

The information contained herein is, to the best of our knowledge and belief, accurate. However, since conditions of handling and use are beyond our control, we do not guarantee any results, and we are not liable for any damage incurred by following these suggestions. Nothing contained herein is to be construed as a recommendation for use in violation of any patent or applicable laws or regulations.